



CERTIFICATE OF ANALYSIS
 Complies with ISO Guide 34, ISO Guide 31,
 ISO Guide 35, and ISO 9001
TRACEABLE® CERTIFIED REFERENCE MATERIAL



This certificate indicates traceability to standards provided by (NIST) National Institute of Standards and Technology and/or a National Standards Laboratory.

Amended Certificate
Reference Certificate 4179-8068866

Certificate No.: 4179-8151705

Description: Conductivity Standard 99,958 µS/cm

Catalog Number: 4179 **Lot:** CC15426

Certificate Date: October 26, 2016 **Expiration Date:** October 26, 2017

Certified Value: 99,958 µS/cm U = ±370 µS/cm (k=2) at 25°C

Derived Values: 99958 micromho/cm, 10.00 ohm-cm, 66639 PPM D.S.

Certification measurements are performed under ISO Guide 34, A2LA accreditation no. 1750.02 and are traceable to recognized national and international standards via an unbroken chain of comparisons. Electrical conductance is the reciprocal of electrical impedance. The International System of units (SI), derived unit of conductance, is Siemens (S), also referred to as (mhos) the reciprocal of ohms. The certified value is expressed in microsiemens per centimeter (µS/cm).

MEASUREMENT: Ten (10) 100 ml samples were measured from this lot. The conductivity of each sample was derived from a measurement of the impedance of the solution using a conductivity meter and calibrated cell. The cell and sample were temperature controlled by submersion in a water bath at 25°C ± 0.015°C.

UNCERTAINTY: The certified value is given as the average of the measured samples. The reported expanded uncertainty (U) is determined from the measurement variation from sample to sample, change due to shelf life, and from the uncertainty of the measurement process. The value of uncertainty is multiplied by k=2, which for a normal distribution corresponds to a coverage probability of approximately 95%. Uncertainty is calculated in accordance with the ISO "Guide to the Expression of Uncertainty in Measurement" (GUM).

METHOD: The certified reference material is prepared and analyzed by Control Company. The certified reference material consists of a mixture of a dilute solution of less than 8% (by mass) potassium chloride (KCL), of less than 2% (by mass) propanol, and of less than 94% (by mass) deionized water in equilibrium with atmospheric carbon dioxide. Mixing was performed by circulation utilizing a proprietary method.


 Aaron Judice, Technical Manager


 Nicol Rodriguez, Quality Manager

Traceability: Standards and Equipment Used

Description	Serial Number	Due Date	Traceable Reference
Conductivity Probe (4W)/ Meter	10124-F02	3/08/17	TC26-7948973
Digital Thermometer	140073819	2/03/17	4000-7392222
Calibration Bath TC-337	B5C477		

Laboratory environment conditions: 25.0°C 53%RH 1,022mb/hPa

CONTROL COMPANY 12554 Old Galveston RD Suite B230 Webster TX 77598 USA
Tel: (281) 482 1714 Fax: (281) 482 9448 sales@control3.com www.control3.com

Control Company is an ISO Guide 34:2009 Certified Reference Material (CRM) Producer Accredited by American Association for Laboratory Accreditation (A2LA Certificate No. 1750.02). This certificate fulfills the requirements of ISO Guide 31:2000 (Reference Materials - Contents of Certificates and Labels), ISO Guide 34:2009 "Quality System Guidelines for the Production of Reference Materials", and ISO Guide 35:2006 "Certification of Reference Materials - General and Statistical Principles". Control Company is an ISO/IEC 17025:2005 Calibration Laboratory Accredited by American Association for Laboratory Accreditation (A2LA Certificate No. 1750.01). Control Company is ISO 9001:2008 certified by DNV GL (Certificate No. CERT-01805-2006-AQ-HOU-RvA). TRACEABLE® is a registered trademark of Control 3 Inc.



CERTIFICATE OF ANALYSIS
Complies with ISO Guide 34, ISO Guide 31,
ISO Guide 35, and ISO 9001
TRACEABLE® CERTIFIED REFERENCE MATERIAL



PACKAGING: This material is available in both a 460 ml bottle and a 100 ml One-Shot™.

INTENDED USE: The certified reference material is intended for the calibration of conductivity cell constants, for conductivity measurement, for the validation of analytical methods, and for the preparation of working reference standards.

INSTRUCTIONS FOR USE: The certified reference material should be open for the minimum time. Rinse the cell in a small amount of the certified reference material and discard. The recommended sample size for measurement is 100 ml. Discard the standard after use and under the following circumstances: if the expiration date is past due, four months after opening, or if any color, turbidity, or visible microbiological growth become evident. Standards which have been opened are not protected from growth. Do not return used solution to this standard. Contaminates and evaporation have a significant effect on conductivity. Keep the standard closed. Keep the standard stored at a stable temperature. Select a standard as near as possible to that of the unknown solution to be measured. Do not standardize at 10,000 uS and then measure unknowns at 100 uS. Reference any accompanying instructions shipped with this product.

Temperature has a significant effect on conductivity. For measurements at a temperature other than 25°C, refer to the temperature correction table provided. This product should be used as near as possible to 25°C.

HOMOGENEITY: Ten (10) 100 ml samples were selected for analytical control. Results from different samples showed no statistically significant differences, nor was there any correlation between values obtained and the bottling sequence. Bottle-to-bottle (One-Shot™ to One-Shot™) variations of the samples measured are included as a part of the calculated measurement uncertainty stated on page 1 of this certificate. A minimum sample size of 100 ml should be used to maintain the certified value and the associated statement of uncertainty. This standard as formulated is considered infinitely soluble.

STABILITY, SHELF LIFE: The expiration date stated on page 1 indicates the period of time which the certified reference material in a properly packaged, unopened, unused, and stored under environmentally controlled and monitored conditions remains within the specified uncertainty range.

EXPIRATION DATE: The date after which a certified reference material should be discarded.

STORAGE: Store below 40°C and above 0°C.

SHIPPING: Ship below 50°C and above 0°C.

MAINTENANCE OF CERTIFICATION: Control Company monitors representative samples from this lot over the period of its certification. If a change occurs that affects the certification before the expiration date, Control Company posts amended certificates at www.control3.com/crmupdate.htm.

MSDS INFORMATION: Please refer to the Material Safety Data sheet for information regarding this certified reference material at www.control3.com (Search MSDS). Use only the first four digits of the certificate number to locate the MSDS.

QUALITY STANDARD DOCUMENTATION:

ISO Guide 34:2009 General Requirements for the Competence of Reference Material Producers, accredited A2LA Certificate Number 1750.02.

ISO Guide 31:2000 Reference Materials - Contents of Certificates and Labels.

ISO Guide 35:2006 Certification of Reference Materials - General and Statistical Principals.

ISO/IEC 17025:2005 General Requirements for the Competence of Testing and Calibration Laboratories, accredited A2LA Certificate Number 1750.01.

ANSI/NCSL Z540-1:1994 Calibration Laboratories and Measuring and Test Equipment-General Requirements.

ISO 9001:2008 Quality Management System Registration - DNV GL Certificate Number CERT-01805-2006-AQ-HOU-RvA

SUPPORTED METHODS: This certified reference material meets test requirements for Federal, State, and local agencies, CAP, CLSI, ACS, and CLIA. Traceable® Certified Reference Material complies with and is essential for use in these official methods: AOAC 973.40, EPA 120.1, Standard Method 2510 (APHA, AWWA, WEF), ISO 7888, DIN 38404, ASTM D1125, USGS I-1780, USP 645, OIML R56, IUPAC, and for A2LA/NVLAP accreditations /ISO 9000 certifications. Material may be used to calibrate all conductivity meters and to determine all conductivity cell constants.

CONTROL COMPANY 12554 Old Galveston RD Suite B230 Webster TX 77598 USA
Tel: (281) 482 1714 Fax: (281) 482 9448 sales@control3.com www.control3.com

Control Company is an ISO Guide 34:2009 Certified Reference Material (CRM) Producer Accredited by American Association for Laboratory Accreditation (A2LA Certificate No. 1750.02). This certificate fulfills the requirements of ISO Guide 31:2000 (Reference Materials - Contents of Certificates and Labels), ISO Guide 34:2009 "Quality System Guidelines for the Production of Reference Materials", and ISO Guide 35:2006 "Certification of Reference Materials - General and Statistical Principals". Control Company is an ISO/IEC 17025:2005 Calibration Laboratory Accredited by American Association for Laboratory Accreditation (A2LA Certificate No. 1750.01). Control Company is ISO 9001:2008 certified by DNV GL (Certificate No. CERT-01805-2006-AQ-HOU-RvA). TRACEABLE® is a registered trademark of Control 3 Inc.



CERTIFICATE OF ANALYSIS
 Complies with ISO Guide 34, ISO Guide 31,
 ISO Guide 35, and ISO 9001
TRACEABLE® CERTIFIED REFERENCE MATERIAL



Temperature Correction Information 1.780 %

If your conductivity meter allows you to set a temperature coefficient (temperature correction) then the underlined number shown above is the best approximation for this specific analysis for this specific Traceable® Certified Reference Material. For more precise measurements use the chart. Use the chart below only for making absolute measurements. That is, measurements without any automatic temperature correction (temperature coefficient set to 0). The chart below displays derived values.

Using a thermometer, measure the temperature of this certified reference material. Shown on the chart is temperature (in the far left column) in whole degrees. Shown across the top row is temperature in tenths of a degree. Locate the measured temperature in whole numbers on the far left column, and then follow across the row to the temperature in tenths of a degree. At the intersection is the certified reference material value at that specific temperature. Standardize your meter using that value. Example: Measured temperature is 25.0°C. Find 25°C in the far left column, find the row 0.0°C. Where 25°C and 0.0°C intersect, read the value in microsiemens/cm.

Temperature Correction Chart in microsiemens/cm

	0.0°C	0.1°C	0.2°C	0.3°C	0.4°C	0.5°C	0.6°C	0.7°C	0.8°C	0.9°C
02°C	60637	60800	60962	61125	61287	61450	61613	61776	61938	62101
03°C	62264	62427	62591	62754	62917	63081	63244	63408	63571	63735
04°C	63899	64063	64227	64391	64555	64719	64883	65047	65212	65376
05°C	65541	65705	65870	66035	66199	66364	66529	66694	66859	67025
06°C	67190	67355	67521	67686	67852	68017	68183	68349	68515	68681
07°C	68847	69013	69179	69345	69511	69678	69844	70011	70177	70344
08°C	70511	70678	70845	71012	71179	71346	71513	71680	71848	72015
09°C	72183	72350	72518	72686	72853	73021	73189	73357	73525	73693
10°C	73862	74030	74198	74367	74535	74704	74873	75041	75210	75379
11°C	75548	75717	75886	76056	76225	76394	76564	76733	76903	77072
12°C	77242	77412	77582	77752	77922	78092	78262	78432	78603	78773
13°C	78944	79114	79285	79456	79626	79797	79968	80139	80310	80481
14°C	80653	80824	80995	81167	81338	81510	81682	81853	82025	82197
15°C	82369	82541	82713	82885	83058	83230	83402	83575	83747	83920
16°C	84093	84266	84438	84611	84784	84957	85131	85304	85477	85651
17°C	85824	85998	86171	86345	86519	86692	86866	87040	87214	87389
18°C	87563	87737	87911	88086	88260	88435	88610	88784	88959	89134
19°C	89309	89484	89659	89834	90009	90185	90360	90536	90711	90887
20°C	91063	91238	91414	91590	91766	91942	92118	92295	92471	92647
21°C	92824	93000	93177	93353	93530	93707	93884	94061	94238	94415
22°C	94592	94769	94947	95124	95302	95479	95657	95835	96012	96190
23°C	96368	96546	96724	96902	97081	97259	97437	97616	97794	97973
24°C	98152	98330	98509	98688	98867	99046	99225	99404	99584	99763
25°C	99958	100122	100301	100481	100661	100841	101020	101200	101380	101561
26°C	101741	101921	102101	102282	102462	102643	102823	103004	103185	103366
27°C	103546	103727	103909	104090	104271	104452	104634	104815	104996	105178
28°C	105360	105541	105723	105905	106087	106269	106451	106633	106816	106998
29°C	107180	107363	107545	107728	107911	108094	108276	108459	108642	108825
30°C	109008	109192	109375	109558	109742	109925	110109	110293	110476	110660
31°C	110844	111028	111212	111396	111580	111765	111949	112133	112318	112502
32°C	112687	112872	113057	113241	113426	113611	113796	113982	114167	114352
33°C	114538	114723	114909	115094	115280	115466	115651	115837	116023	116209
34°C	116396	116582	116768	116954	117141	117327	117514	117700	117887	118074
35°C	118261	118448	118635	118822	119009	119196	119384	119571	119759	119946

CONTROL COMPANY 12554 Old Galveston RD Suite B230 Webster TX 77598 USA
Tel: (281) 482 1714 Fax: (281) 482 9448 sales@control3.com www.control3.com

Control Company is an ISO Guide 34:2009 Certified Reference Material (CRM) Producer Accredited by American Association for Laboratory Accreditation (A2LA Certificate No. 1750.02). This certificate fulfills the requirements of ISO Guide 31:2000 (Reference Materials - Contents of Certificates and Labels), ISO Guide 34:2009 "Quality System Guidelines for the Production of Reference Materials", and ISO Guide 35:2006 "Certification of Reference Materials - General and Statistical Principles". Control Company is an ISO/IEC 17025:2005 Calibration Laboratory Accredited by American Association for Laboratory Accreditation (A2LA Certificate No. 1750.01). Control Company is ISO 9001:2008 certified by DNV GL (Certificate No. CERT-01805-2006-AQ-HOU-RvA). TRACEABLE® is a registered trademark of Control 3 Inc.