



SCOPE OF ACCREDITATION TO ISO 17034:2016

CONTROL COMPANY
12554 Galveston Road Suite B230
Webster, TX 77598
Marisa Elms Phone: 281 400 2104

REFERENCE MATERIAL PRODUCER

Valid To: December 31, 2021

Certificate Number: 1750.02

In recognition of the successful completion of the A2LA evaluation process, accreditation is granted to this Reference Material Producer for the production of reference materials and certified reference materials of the following categories:

Category and sub-category of Certified Reference Material	Concentration Ranges	Uncertainty	Measurement Technique(s)
Category A9: Conductivity Standards	(1 to 10) $\mu\text{S/cm}$ (50 to 100) $\mu\text{S/cm}$ (1000 to 1413) $\mu\text{S/cm}$ (5000 to 10 000) $\mu\text{S/cm}$ (50 000 to 100 000) $\mu\text{S/cm}$ (150 000 to 200 000) $\mu\text{S/cm}$	0.62 $\mu\text{S/cm}$ 2.1 $\mu\text{S/cm}$ 4.6 $\mu\text{S/cm}$ 40 $\mu\text{S/cm}$ 360 $\mu\text{S/cm}$ 510 $\mu\text{S/cm}$	Comparison to NMI SRM
pH Buffers	(4, 7, 10) pH	0.011 pH	Comparison to NMI SRM

Category and sub-category of Reference Material	Concentration Ranges	Methods
Category A9: Conductivity Standards	(1 to 200 000) $\mu\text{S/cm}$	Testing with calibrated probe
pH Buffers	(1 to 12) pH	Testing with calibrated probe