



SCOPE OF ACCREDITATION TO ISO GUIDE 34:2009

CONTROL COMPANY
 12554 Galveston Road Suite B230
 Webster, TX 77598
 Aaron Judice Phone: 281 482 1714

REFERENCE MATERIAL PRODUCER¹

Valid To: December 31, 2017

Certificate Number: 1750.02

In recognition of the successful completion of the A2LA evaluation process, accreditation is granted to this Reference Material Producer for the production of certified reference materials of the following category:

Standards	Concentration Ranges	Uncertainty	Methods
Category A9: Conductivity Standards	(1 to 10) $\mu\text{S/cm}$ (50 to 100) $\mu\text{S/cm}$ (1000 to 1413) $\mu\text{S/cm}$ (5000 to 10 000) $\mu\text{S/cm}$ (50 000 to 100 000) $\mu\text{S/cm}$ (150 000 to 200 000) $\mu\text{S/cm}$	0.55 $\mu\text{S/cm}$ 2.1 $\mu\text{S/cm}$ 4.6 $\mu\text{S/cm}$ 40 $\mu\text{S/cm}$ 360 $\mu\text{S/cm}$ 510 $\mu\text{S/cm}$	Comparison to NMI SRM
pH Buffers	(4, 7, 10) pH	0.011 pH	Comparison to NMI SRM

¹ This reference material producer is approved to produce Certified Reference Materials (CRM) for all items listed on the scope of accreditation.